

**Notice**  
**Subj / Inv.67/2021/2**  
**(General Invitation)**  
**Supplying Cathlab single plane & Biplane**

Dear Sirs;

We would like to attract your attention that the invitation has been amended as following:

- 1- Article (second party adheres to use the technique that reduce the radiation dose to the lowest possible extent for CT SCAN contracts) has been deleted from technical specifications.
- 2- Following two articles to be added

**Please take action in this respect**

**B.RGDS**

---

تنويه  
م/ المفاتحة المرقمة ٦٧/٢٠٢١/٢  
دعوة عامة  
( تجهيز اجهزة القسطرة القلبية احادي الذراع وثنائي الذراع )

**تحية طبية**

نود ان نبين أنه تم تعديل المفاتحة كالاتي :-

- ١- تم حذف الفقرة (١) - يلتزم الطرف الثاني بالنسبة لعقود CT Scan استخدام التقنية التي تساهم بتقليل جرعة الاشعاع الى ادنى حد ممكن ( من المواصفات الفنية .
- ٢- يتم اضافة الفقرتين ادناه

<b>Training</b>	<ul style="list-style-type: none"> <li>- Local training for <math>\geq 2</math> physician, 2 operators to operate the equipment ,1 Eng. for application (as clinical Eng.), 2 Eng. for maintenance following and simple maintenance for 1 weeks</li> <li>- Abroad training <math>\geq 1</math> week in specialized training center for 1 physician, 1 operator for operate the equipment , 1 Eng. (from medical equipment department- technical affairs directorate). 1 Eng. for maintenance following and simple maintenance.</li> </ul>
<b>POWER REQUIREMENTS</b> Line power UPS for whole system	380-480 VAC, 50/60 Hz, 3-phase Yes $\geq 120$ KVA

للتفضل بالأطلاع واتخاذ ما يلزم .... مع الأ احترام